


ACCESSORIES

SHUNT CAL / RCAL RESISTORS

Drawing Number: FI1208-A


POWER RATING	MATERIAL
1/4 WATT	PRECISION METAL FILM
TOLERANCE	TEMPERATURE COEFFICIENT
0.01 %	10 PPM



STOCK#	SHUNT CAL VALUE	BRIDGE RESISTANCE	SHUNTED OUTPUT
GOD01353	43.575K OHMS	350 OHMS	2.00 mV/V
GOD01354	60.4K OHMS	350 OHMS	1.44 mV/V
GOD01355	87.325K OHMS	700 OHMS	2.00 mV/V
GOD01356	100K OHMS	350 OHMS	0.87 mV/V
		700 OHMS	1.74 mV/V
		1000 OHMS	2.49 mV/V
GOD01357	150K OHMS	700 OHMS	1.16 mV/V
		1000 OHMS	1.66 mV/V
GOD01358	174.825K OHMS	350 OHMS	0.50 mV/V
GOD01359	200K OHMS	700 OHMS	0.87 mV/V
		1000 OHMS	1.25 mV/V
GOD01360	301K OHMS	1000 OHMS	0.83 mV/V

SHUNT CALIBRATION:

ELECTRICAL SIMULATION OF TRANSDUCER OUTPUT BY INSERTION OF KNOWN SHUNT RESISTORS BETWEEN APPROPRIATE POINTS WITHIN THE CIRCUITRY



FOR POSITIVE OUTPUT DURING SHUNT CALIBRATION,
RESISTOR MUST BE CONNECTED ACROSS -EXCITATION AND -SIGNAL

SCIENCE GATE
Your Automation Partner



SCIGATE AUTOMATION (S) PTE LTD

No.1 Bukit Batok Street 22 #01-01 Singapore 659592

Tel: (65) 6561 0488

Fax: (65) 6562 0588

Email: sales@scigate.com.sg

Web: www.scigate.com.sg

Business Hours: Monday - Friday 8.30am - 6.15pm

FUTEK
ADVANCED SENSOR TECHNOLOGY, INC.

This drawing is submitted solely for the information and exclusive use of the original addressee. It is not to be divulged in whole or in part, by any firm or individual without written permission from FUTEK

10 THOMAS
IRVINE, CA 92618 USA
1-800-23-FUTEK (38835)

INTERNET:
<http://www.futek.com>